

## UKCA Declaration of Conformity

For the product described as follows

Product designation

**Diaphragm pressure gauge**  
(with contact desvice KE ## S/M ##### H2)

Type designation

**MA15 ... 10**

is hereby declared to comply with the essential requirements, specified in the following UK regulations:

**Statutory regulation No.**

2016 No. 1101

2022 No. 1647

2021 No. 422

**Description**

*The Electrical Equipment (Safety) Regulations 2016*

*The Hazardous Substances and Packaging (Legislative Functions and Amendment) (EU Exit) Regulations 2020*

*The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment (Amendment) Regulations 2021*

The products have been tested according to the following standards.

**Low Voltage Directive (LVD):**

*BS EN 61010-1+A1:2017-03-31*

*Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements*

**Restriction of Hazardous Substances (RoHS):**

*BS EN IEC 63000:2018-12-10*

*Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances*

The sole responsibility for drawing up this declaration of conformity in relation to the fulfilment of the essential requirements and the preparation of the technical documentation lies with the manufacturer.

**Manufacturer**

**FISCHER Mess- und Regeltechnik GmbH**

Bielefelder Str. 37a  
32107 Bad Salzufflen, Germany

Tel. +49 (0)5222 974 0

The devices bear  
the following marking:

**UK  
CA**



Bad Salzufflen  
04 Okt 2021

G. Gödde  
Managing director

